

MR-Linac Level 2 (Zone IV) Personnel Safety Competency Evaluation Form

Staff: _____ Date: _____

#	Description of Competency	Met	Not Met
0	Candidate has documented annual MRI safety training		
	Identification of MRI Safety Zones		
1a	Candidate is able to identify MRI Zone III boundaries in MR-Linac suite and articulate what staff are allowed to work in Zone III.		
1b	Candidate is able to identify MRI Zone IV boundaries in MR-Linac suite and articulate qualifications of staff allowed to work in Zone IV.		
	Identification of MRI Safety Equipment Labels		
2a	Candidate is able to identify MRI Safe equipment label and articulate how the equipment can be used in Zone IV.		
2b	Candidate is able to identify MRI Conditional equipment label and articulate how the equipment can be used in Zone IV.		
2c	Candidate is able to identify MRI Unsafe equipment label and articulate that equipment cannot enter Zone IV.		
2d	Candidate is able to articulate what process to follow with unlabeled equipment.		
	Identification of MRI Safety Equipment		
3a	Candidate is able to locate MRI screening forms and articulate what should be done when implant is detected.		
3b	Candidate is able to locate ferromagnetic guards in Zone III and articulate what should be done if guard alarm is tripped.		
3c	Candidate is able to locate and use portable ferromagnetic wand, and articulate what should be done if wand alarm is tripped.		
3d	Candidate is able to locate MRI-safe emergency stretcher and slider board.		
3e	Candidate is able to locate emergency help buttons in Zone III and Zone IV.		
3f	Candidate is able to locate MRI-conditional fire extinguisher.		
3g	Candidate is able to locate electrical emergency-off buttons in Zone III and Zone IV and articulate when they are to be used.		
3h	Candidate is able to locate emergency quench buttons in Zone III and Zone IV and articulate when they are to be used.		
3i	Candidate is able to locate manual vault door crank.		
	Handling of Unresponsive Patient or Volunteer		
4a	Candidate is able to articulate the process of handling a patient or volunteer emergency in the MR-Linac suite.		
4b	Candidate is able to retrieve and setup MRI-safe emergency stretcher.		
4c	Candidate is able to transfer volunteer from MR-Linac table to MRI-safe emergency stretcher.		
	Fire Alarm Activation		
5	Candidate is able to articulate procedure to follow in the event of a fire in Zones III or IV.		
	Loss of Electrical Power		
6	Candidate is able to manually open vault door using crank.		

Evaluator: _____ Date: _____